

# Material: ASTM F 2527 R30003

## Standard Specification for Seamless Drawn Cobalt Alloy Small Diameter Tubing for Surgical Implants

**Group:** Non-Ferrous Nickel Alloys

**Sub Group:** ASTM F 2527 R30003 Seamless Drawn Cobalt Alloy Small Diameter Tubing for Surgical Implants

**Application:** Intended for Valve, Pump, General Engineering, Automotive, Medical and other Industries

**Grade Belongs to the Industry:** Casting

Chemical Composition		
Carbon	C %	0.150 max.
Chromium	Cr %	19.000 - 21.000
Molybdenum	Mo %	6.000 - 8.000
Nickel	Ni %	14.000 - 18.000
Cobalt	Co %	39.000 - 41.000
Silicon	Si %	1.200 max.
Manganese	Mn %	1.500 - 2.500
Sulphur	S %	0.015 max.
Phosphorus	P %	0.015 max.
Beryllium	Be %	0.100 max.
Iron	Fe %	Balance
-	-	-
-	-	-
-	-	-
-	-	-

Heat Treatment	
	As-Cast or Annealing

Mechanical Properties	
Tensile Strength in Mpa	850 min.
Yield Strength in Mpa	450 min.
Elongation in %	1 min.
Reduction of Area in %	-
Hardness in BHN	-
Impact in Joule	-

Cross Reference Table			
Material	Standard	Country	Grade Belong to the Industry
F1058 Grade 1	ASTM	USA	Wire and Strip
R30008	UNS	USA	Casting
7402-4	JIS	Japan	Casting
-	-	-	-
-	-	-	-
-	-	-	-
-	-	-	-

**Disclaimer:** All information displayed in our data sheets are for reference purpose only and are sole property of their respective owners. Information and or material are used for educational purposes only. Data at actual may vary at actual and case to case basis. ICAST Alloys LLP does not guarantee validity of these parameters. Warranties and liabilities are exclusive to our terms and conditions of business.

Customer Care: +91-99090 45075 Email: [info@icastllp.com](mailto:info@icastllp.com)



+91-99090 45075



[info@icastllp.com](mailto:info@icastllp.com)



ICAST ALLOYS LLP, Plot 2527, Road H1, Kranti Gate, GIDC Metoda, Lodhika, Rajkot-360021, Gujarat, India